



Texas Department of State Health Services

PO BOX 149347
 AUSTIN, TEXAS 78714-9347
 1-888-963-7111
 www.dshs.state.tx.us

LABORATORY SERVICES SECTION, MC-1947
 1100 W. 49th St., Austin, Tx. 78756 (512)458-7587

Semivolatiles Organic Analysis Report

Submitter Identification Number: 2270091

HAZY HILLS WSC
 THERING, MIKE
 22306 HAZY HOLLOW DR UNIT 3
 SPICEWOOD, TX 78669-6699

Date Reported : 07/18/2023
 Report ID# : 20230718121534AG41751

Lab Sample ID# : AG41751	Water Source :	Date Collected : 05/23/2023 14:43	Conc. Units : µg/L
Sample Priority : NORMAL	Entry Point(s) : EP001	Date Received : 05/24/2023	Method : EPA 525.2
TCEQ ID#(s) : 2308535		Date Analyzed : 06/02/2023	Analyst : RR
		Extraction Date : 05/31/2023	Sample Cond. : Acceptable

Regulated Compounds	Result	Qualifier	Monitored Compounds continued	Result	Qualifier
Alachlor ¹	<0.2		Dimethylphthalate	<2.0	
Atrazine ¹	<0.1		Fluorene	<0.20	
Benzo[a]pyrene ¹	<0.02		2,2',3,3',4,4',6-Heptachlorobiphenyl	<0.51	
alpha-Chlordane	<0.2		2,2',4,4',5,6'-Hexachlorobiphenyl	<0.20	
gamma-Chlordane	<0.2		Indeno[1,2,3-cd]pyrene	<0.20	
trans-Nonachlor	<0.2		Metolachlor	<0.20	
Di(2-ethylhexyl) adipate ¹	<0.6		Metribuzin	<0.20	
Di(2-ethylhexyl) phthalate ¹	<0.6		Naphthalene	<0.20	
Heptachlor ¹	<0.04		2,2',3,3',4,5',6,6'-Octachlorobiphenyl	<0.51	
Hexachlorobenzene ¹	<0.1		2,2',3',4,6-Pentachlorobiphenyl	<0.20	
Hexachlorocyclopentadiene ¹	<0.1	*	Phenanthrene	<0.20	
Lindane ¹	<0.02		Propachlor	<0.20	
Methoxychlor ¹	<0.1		Pyrene	<0.20	
Simazine ¹	<0.07		2,2',4,4'-Tetrachlorobiphenyl	<0.20	
			2,4,5-Trichlorobiphenyl	<0.20	
Monitored Compounds	Result	Qualifier	Trifluralin	<0.20	
Acenaphthene	<0.20				
Acenaphthylene	<0.20				
Aldrin	<0.20	*			
Anthracene	<0.20				
Benzo(a)anthracene	<0.20				
Benzo[b]fluoranthene	<0.20				
Benzo[g,h,i]perylene	<0.20				
Benzo[k]fluoranthene	<0.20				
Bromacil	<0.20	K			
Butachlor	<0.20				
Butylbenzylphthalate	<2.0				
2-Chlorobiphenyl	<0.20				
Chrysene	<0.20				
Dibenz[a,h]anthracene	<0.20				
Di-n-butylphthalate	<2.0				
2,3-Dichlorobiphenyl	<0.20				
Dieldrin	<0.20				
Diethylphthalate	<2.0				

Comments:

* - This analyte has known instability and/or method performance issues and quantitation should be considered approximate.
 K - The associated laboratory fortified blank spike (and/or its duplicate) recovery was above method acceptance limits.
 1 of 3 EPA method 525.2 internal standard was below QC limits in AG41751. The test results on this report relate only to the sample identified on this report. The test results for analytes noted(1) meet all TNI (2016 Standard) requirements.

Authorized by Group Manager AVINYARD on 07/13/2023



Texas Department of State Health Services

LABORATORY SERVICES SECTION, MC-1947
1100 W. 49th St., Austin, Tx. 78756 (512)458-7587

PO BOX 149347
AUSTIN, TEXAS 78714-9347
1-888-963-7111
www.dshs.state.tx.us

Pesticides by Method 508.1 Analysis Report

Submitter Identification Number: 2270091

HAZY HILLS WSC
THERING, MIKE
22306 HAZY HOLLOW DR UNIT 3
SPICEWOOD, TX 78669-6699

Date Reported : 07/18/2023
Report ID# : 20230718121534AG41751

Lab Sample ID# : AG41751 Water Source :
Sample Priority : NORMAL Entry Point(s) : EP001
TCEQ ID#(s) : 2308535

Date Collected : 05/23/2023 14:43 Conc. Units : ug/L
Date Received : 05/24/2023 Method : 508.1 Rev. 2.0
Date Analyzed : 06/07/2023 Analyst : JH
Sample Cond. : Acceptable

Regulated Compounds	Result	Qualifier
Chlordane ¹	<0.2	
Endrin ¹	<0.01	
Heptachlor epoxide ¹	<0.02	
Toxaphene ¹	<1.	
Screened Compounds	Result	Qualifier
Aroclor 1016 ²	<0.08	
Aroclor 1221 ²	<20.	
Aroclor 1232 ²	<0.5	
Aroclor 1242 ²	<0.3	
Aroclor 1248 ²	<0.1	
Aroclor 1254 ²	<0.1	
Aroclor 1260 ²	<0.2	

Comments:

1 of 3 EPA method 525.2 internal standard was below QC limits in AG41751. The test results on this report relate only to the sample identified on this report. The test results for analytes noted⁽¹⁾ meet all TNI (2016 Standard) requirements. The test results for analytes noted⁽²⁾ meet all TNI (2016 Standard) requirements for Aroclor Identification. Aroclor quantitation is not accredited.

Authorized by Group Manager AVINYARD on 07/13/2023